## ahn my Lab CLC-01

## Clinical Centrifuge 4000 rpm

Usage - designed for blood centrifugation and other clinical applications

Multiple rotor and tube options - 6x10 mL (swing out rotor) /8x15 mL (fixed angle rotor) / 16x10 mL (fixed angle rotor) / with each rotor, the appropriate tube adaptors, sleeves, spacers are included - 10 mL, 9 mL, 8 mL, 6 mL, 4 mL, 3 mL and 2 mL

Maximum Speed - up to 4000 rpm / 2270 x g varies per rotor Precise operation - speed interval of 10 rpm starting from 500 rpm

Memory feature - can save upto 99 user defined programs (protocols)

Increased user safety and comfort - automatic safety brake system preventing the lid from opening upon operation / imbalance detection

**Robust design** - brushless DC motor enables maintenance free and long lasting operation, perfect for extended runs

**Comfortable use** - last run memory function / user-friendly interface for speed and time setting / quick spin feature



## **Ordering information**

AHN mvLab® CLC-01 Clinical Centrifuge 4000 rpm

## **Specifications** Rotor Type: **Rotor Type: Rotor Type:** Fixed angle Fixed angle Swing out Capacity: 6x10 mL Capacity: 8x15 mL Capacity: 16x10 mL Max RCF: 2270 x g Max RCF: 1950 x g Max RCF: 2180 x g Cat. No.: Cat. No.: Cat. No.: CRC-4-6S CRC-4-8 CRC-4-16

Cat. No.

7-010-00-0

71 II VIII JEAN OLO OT CIII II CAI	7 010 00 0
Specifications	//
Imbalance detection	yes, auto cut off
Timer setting	1-999 mins / infinite mode
Weight	7.7 kg (without rotor)
Dimensions (WxDxH)	355x415x173

Delivery Package	
1 pc. Clinical centrifuge 4000 rpm	
1 pc. Power supply adaptor	
1 pc. T-Allen Key	
Reduction adaptors	
1 pc. Tube holder	
1 pc. Instruction manual	
1 pc. Warranty card	

AHN myLab CLC-01 Clinical Centrifuge 4000 rpm, is designed for processing blood and other clinical samples in applications like PRP, medical practitioners, pathological laboratories, as well as for routine hospital centrifuging test. In particular it is used for processing and analyzing samples from human body in in vitro diagnostic applications to ensure that in vitro diagnostic device can be used according to Directive 98/79/EC. Therefore AHN myLab CLC-01 Clinical Centrifuge 4000 rpm, and its accessories being in-vitro diagnostic accessories are themselves in-vitro diagnostic medical devices within the meaning of directive 98/79/EC of the European Parliament and of the Council of 27 October 1009 on its vitro diagnostic medical devices. 27 October 1998 on in-vitro diagnostic medical devices



**Description** 















